



**ChemPure* Brand
Chemicals**

...Pure Quality

Date : 11/06/2013

CERTIFICATE OF ANALYSIS

PRODUCT/PRODUIT: CP-A1238, -C1238, -E1238, -U1238

Buffer Solution

Tampon de Référence

Yellow pH 7.00 Jaune

LOT: J1613 (B5205)

TEST

Phosphate Buffer

Traceable to N.I.S.T.

pH @ 25°C

Tampon de Phosphate

Conforme aux normes de N.I.S.T.

SPECIFICATIONS

7.00 ± 0.01

RESULTS/ RESULTATS

7.01

FW/PM

VRM

Vyto Mekesa

Quality Control Department

Dépt. Contrôle de la Qualité