



**ChemPure* Brand
Chemicals**

...Pure Quality

April, 30 2014

CERTIFICATE OF ANALYSIS

Product/Produit: CP-A1238P, C1238, E1238

BUFFER SOLUTION

Green pH 7.00 Vert

LOT: B5200-K2713

TEST:

SPECIFICATIONS

RESULTS/ RESULTATS

Phosphate Buffer

Traceable to N.I.S.T.

pH @ 25°C

Tampon de Phosphate

Conforme aux normes de N.I.S.T.

7.00 ± 0.01

7.00

vkm

Quality Control Department

Dépt. Contrôle de la Qualité