

Certificate of Analysis

Hydrochloric Acid, 1.00 Normal

Manufacture Date: 01/28/2016

Expiration Date: 12/05/2017

Lot Number: 4601E00 Product Number: CP-2412

The certified value for this product is confirmed in independent testing by a second qualified chemist.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Hydrochloric Acid	7647-01-0	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs.	0.999-1.001 N at 20°C	1.000 N at 20°C	723
Tris(hydroxymethyl)aminomethane/Methyl Red)			

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part NumberSize / Package TypeShelf Life (Unopened Container)CP-A2412P500 mL natural poly24 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Katie Schnur

Quality Control Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

Version: 1.3 Lot Number: 4601E00 Product Number: CP-2412 Page 1 of 1