



...Pure Quality

---

## Certificate of Analysis

---

**Product:** ACS/HPLC Acetonitrile  
**Cat:** CP-C0157  
**Lot:** DT990X  
**Date of Manufacture:** 3/23/18  
**Use Before Date:** 3/22/21

Meets ACS Specifications

| <u>Test</u>                    | <u>Specification</u> |             | <u>Result</u> | <u>Units</u> |
|--------------------------------|----------------------|-------------|---------------|--------------|
|                                | <u>Min.</u>          | <u>Max.</u> |               |              |
| Appearance                     | Clear                |             | Clear         |              |
| Water by Karl Fisher Titration |                      | 0.05        | <0.001        | %            |
| UV Cutoff                      |                      | 190         | 188           | nm           |
| Refractive Index (20°C)        | 1.3434               | 1.3446      | 1.3438        |              |
| Residue                        |                      | 3           | <1            | mg/L         |
| GC Analysis                    | 99.9                 |             | 99.99         | %            |
| Color (APHA)                   |                      | 10          | <10           |              |
| Titration Acid                 |                      | 8.0         | 0.35          | µeq/g        |
| Titration Base                 |                      | 0.6         | 0.01          | µeq/g        |
| HPLC Gradient @ 254 nm         |                      | 0.005       | <0.0050       | AU           |
| UV Absorbance @ 200 nm         |                      | 0.070       | 0.0418        | AU           |
| UV Absorbance @ 210 nm         |                      | 0.040       | 0.0038        | AU           |
| UV Absorbance @ 225 nm         |                      | 0.020       | <0.0001       | AU           |
| UV Absorbance @ 250 nm         |                      | 0.005       | <0.0001       | AU           |
| UV Absorbance @ 350 nm         |                      | 0.005       | <0.0001       | AU           |

Quality Control Approval

*Janna Dickinson*

LIMS Sample No.: AJ00920  
3/27/2018