



ChemPure Brand Chemicals  
PO Box 5483  
Plymouth, MI 48170  
734-729-1805

# Certificate of Analysis

**Buffer, Reference Standard, pH 4.00 ± 0.01 at 25°C (Color Coded Red)**

**Lot Number:** 4805H54

**Product Number:** CP-1228

**Manufacture Date:** 05/17/2018

**Expiration Date:** 05/05/2020

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST Traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C	0	5	10	15	20	25	30	35	40	45	50
pH	4.00	4.00	4.00	4.00	4.00	4.00	4.01	4.02	4.03	4.04	4.06

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Acid Phthalate	877-24-7	Buffer
Preservative	Proprietary	Commercial
Red Dye	Proprietary	Purified

Test	Specification	Result	NIST SRM#
Appearance	Red liquid	Passed	
pH at 25°C (Method: SQCP027)	3.990-4.010	4.005	185i, 186-I-g, 186-II-g
pH at 25°C (Method: SQCP027) Uncertainty	0.01	0.01	

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (L-A-B Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution.

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
CP-A1228P	500 mL natural poly	24 months
CP-C1228P	4 L natural poly	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

Jim Gibbs

Quality Control Supervisor

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."



**LABORATORY  
ACCREDITATION  
BUREAU** a division of A-S-B  
**ACCREDITED** ISO/IEC 17025  
Certificate # L2387.02 Testing

This test report shall not be reproduced, except in full, without the written approval of **ChemPure Brand Chemicals**.