



**ChemPure* Brand
Chemicals**

...Pure Quality

Date : 15/05/2013

CERTIFICATE OF ANALYSIS

PRODUCT/PRODUIT: CP-A1228, -C1228, -E1228, -U1228

Buffer Standard

Tampon de Référence

Red pH 4.00 Rouge

LOT: A3113 (B5140)

TEST

Potassium Biphthalate Buffer

Traceable to N.I.S.T.

pH @ 25°C

Tampon de Biphthalate de Potassium

Conforme aux normes de N.I.S.T.

SPECIFICATIONS

4.00 ± 0.01

RESULTS/ RESULTATS

4.00

FW/PM

VKM

Quality Control Department

Dépt. Contrôle de la Qualité