



**ChemPure\* Brand  
Chemicals**

...Pure Quality

Date : 06/11/2015

## CERTIFICATE OF ANALYSIS

PRODUCT/PRODUIT: CP-A1238(P), -C1238, -E1238, -U1238

Buffer Solution

Tampon de Référence

Yellow pH 7.00 Jaune

LOT: F1115

FW/PM:

Manufactured Date: June 11, 2015

Expiration: June 11, 2017

### TEST

Phosphate Buffer

Traceable to N.I.S.T.

pH @ 15°C

pH @ 20°C

pH @ 25°C

pH @ 26°C

pH @ 30°C

pH @ 35°C

pH @ 40°C

Tampon de Phosphate Conforme aux normes de N.I.S.T.

### SPECIFICATIONS

7.00 ± 0.01

### RESULTS/RESULTATS

7.04

7.02

7.00

7.00

6.99

6.98

6.98

*VRM*

Vyto Mekesa

Quality Control Department

Dépt. Contrôle de la Qualité