



**ChemPure* Brand
Chemicals**

...Pure Quality

Date : 12/30/2015

CERTIFICATE OF ANALYSIS

PRODUCT/PRODUIT: CP-A1248, -C1248, -E1248, -U1248

Buffer Solution

Tampon de Référence

Blue pH 10.00 Bleu

LOT: G0215 (B5260)

TEST

Borate Buffer

Traceable to N.I.S.T.

pH @ 25°C

Tampon de Borate

Conforme aux normes de N.I.S.T.

SPECIFICATIONS

10.00 ± 0.01

RESULTS/ RESULTATS

10.00

FW/PM

VKM

Quality Control Department

Dépt. Contrôle de la Qualité