



ChemPure Brand Chemicals
39103 Warren Road
Westland, MI 48185
PH: 734-729-1805
support@chempurebrand.com

Certificate of Analysis

Water, ACS Reagent Grade, ASTM Type I, ASTM Type II

Lot Number: 4004L27

Product Number: CP-0204

Date Packaged: 04/24/2020

Expiration Date: 10/13/2021

This product is manufactured by a continuous process that has been validated to produce high purity Water meeting, as a minimum, the specifications below at the time of packaging. This water also meets ASTM Type III and Type IV specifications. The analysis results are from the most recent scheduled monitoring.

This product is not certified to meet USP and EP Purified Water specifications.

Contains

H₂O CAS No. 7732-18-5 MW 18.02

Test	Specification	Result
Appearance	Colorless liquid	Passed
Chloride (Cl)	max 1 ppb	< 0.18 ppb
Conductivity at 25°C (Method: SQCP031, SQCP033)	max 0.056 µS/cm	0.055 µS/cm
Heavy Metals (as Pb)	max 0.01 ppm	< 0.01 ppm
Nitrate (NO ₃)	max 0.4 ppm	< 0.00004 ppm
Organic Carbon (TOC)	max 50 ppb	< 16 ppb
Oxidizable Substances (Permanganate Retention)	To Pass Test	Passed
Phosphate (PO ₄)	max 1.0 ppm	< 0.00002 ppm
Silicate (as SiO ₂)	max 3 ppb	< 1 ppb
Sodium (Na)	max 1 ppb	< 0.09 ppb
Sulfate (SO ₄)	max 1.0 ppm	< 0.00005 ppm

This water is specially purified by conventional water softening, prefiltration, activated carbon organic adsorption, reverse osmosis, mixed bed triple deionization, ultraviolet light irradiation, and 0.2 micron membrane filtration.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
CP-C0204	4 L amber glass	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Jim Gibbs

Quality Control Supervisor

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."