



ChemPure Brand Chemicals  
 39103 Warren Road  
 Westland, MI 48185  
 734-729-1805  
 support@chempurebrand.com

# Certificate of Analysis

**Silver Nitrate, 0.100 Normal (N/10)**

**Lot Number:** 4009B60

**Product Number:** CP-2248

**Manufacture Date:** 09/03/2020

**Expiration Date:** 08/23/2022

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Silver Nitrate	7761-88-8	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Potassium Chloride/Potassium Chromate)	0.0999-0.1001 N at 20°C	0.1000 N at 20°C	999

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
CP-C2248P	4 L black poly	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

Chris Collins  
 Quality Control Supervisor

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."