



Certificate of Analysis

Sulfuric Acid, 1.00 Normal

Lot Number: 4906B43

Product Number: CP-C2485P

Manufacture Date: JUN 11, 2019

Expiration Date: MAY 2022

The certified value for this product is confirmed in independent testing by a second qualified chemist.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sulfuric Acid	7664-93-9	ACS

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Tris(hydroxymethyl)aminomethane/Methyl Red)	0.999-1.001 N at 20°C	1.001 N at 20°C	723

Specification	Reference
Sulfuric Acid, 1 N	APHA (4500-Norg C)
Sulfuric acid, 1.0 N	TAPPI (T 625 cm-85)
Acid Solution, 1 N	APHA (5210 B)
Sulfuric Acid, 1 N	APHA (4500-SiO2 C)
Regenerant Solution 1, H2SO4, 1 N	APHA (4110 B)
Sulfuric Acid, 1 N	APHA (5530 B)
Sulfuric Acid, 1 N	APHA (4500-Cl- B)
Sulfuric Acid, 1 N	APHA (5570 C)
Sulfuric Acid Solution, 1 N	APHA (4500-CN- H)
Sulfuric Acid, 1 N	EPA (425.1)
Sulfuric Acid, 1 N	EPA (SW-846) (9067)
Sulfuric Acid, 1 N	EPA (370.1)
Sulfuric acid, 1 N	TAPPI (T 610 sp-97)
Sulfuric Acid, 1 N	EPA (420.3)
Sulfuric Acid, 1 Normal Volumetric Solution	ACS (N/A)
0.5 M Sulphuric acid Volumetric Solution	EP (3007800)
Sulfuric Acid, Normal (1 N) Volumetric Solution	USP (N/A)
Sulfuric Acid, 1 N	APHA (4500-NH3 B)

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
CP-C2485P	4 L natural poly	36 months
CP-B2485P	500 mL natural poly	36 months
CP-C2485P-1CS	4 x 4 L natural poly	36 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)



Jim Gibbs (06/11/2019)

Quality Control Supervisor

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of ChemPure Brand Chemicals