



Certificate of Analysis

Water, ACS Reagent Grade, ASTM Type I, ASTM Type II

Lot Number: 4103J02

Product Number: CP-0204

Date Packaged: 03/22/2021

Expiration Date: 09/09/2022

This product is manufactured by a continuous process that has been validated to produce high purity Water meeting, as a minimum, the specifications below at the time of packaging. This water also meets ASTM Type III and Type IV specifications. The analysis results are from the most recent scheduled monitoring.

This product is not certified to meet USP and EP Purified Water specifications, inquire if you need this type of water.

Contains

H₂O CAS No. 7732-18-5 MW 18.02

Test	Specification	Result
Appearance	Colorless liquid	Passed
Chloride (Cl)	max 1 ppb	0.08 ppb
Conductivity at 25°C (Method: SQCP031, SQCP033)	max 0.056 µS/cm	0.055 µS/cm
Heavy Metals (as Pb)	max 0.01 ppm	0.01 ppm
Nitrate (NO ₃)	max 0.4 ppm	0.0001 ppm
Organic Carbon (TOC)	max 50 ppb	< 10 ppb
Oxidizable Substances (Permanganate Retention)	To Pass Test	Passed
Phosphate (PO ₄)	max 1.0 ppm	< 0.0001 ppm
Silicate (as SiO ₂)	max 3 ppb	2 ppb
Sodium (Na)	max 1 ppb	0.04 ppb
Sulfate (SO ₄)	max 1.0 ppm	< 0.0001 ppm

This water is specially purified by conventional water softening, prefiltration, activated carbon organic adsorption, reverse osmosis, mixed bed triple deionization, ultraviolet light irradiation, and 0.2 micron membrane filtration. Consult RCC Technical Reference Document TRD02 (for lot numbers starting with 1) or TRD03 (for lot numbers starting with 2) for details of the water purification process.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
CP-E0204C	20 L Cubitainer®	18 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Jon Russell

Quality Control Supervisor

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."