



ChemPure Brand Chemicals  
 PO Box 5483  
 Plymouth, MI 48170  
 734-729-1805

# Certificate of Analysis

## Sodium Hydroxide, 10.0 Normal

Lot Number: 4810J82

Product Number: CP-2393

Manufacture Date: 10/15/2018

Expiration Date: 09/30/2020

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Hydroxide	1310-73-2	Reagent

Test	Specification	Result	NIST SRM#
Appearance	Colorless liquid	Passed	
Assay (vs. Sulfuric Acid/Phenolphthalein)	9.95-10.05 N at 20°C	9.95 N at 20°C	723
Heavy Metals (as Pb)	max 10 ppm	<= 10 ppm	
Iron (Fe)	max 5 ppm	<= 5 ppm	3126

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
CP-E2393	20 L Cubitainer®	24 months

**Recommended Storage:** 15°C - 30°C (59°F - 86°F)

Jim Gibbs  
 Quality Control Supervisor

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This test report shall not be reproduced, except in full, without the written approval of ChemPure Chemicals.