



...Pure Quality

---

## Certificate of Analysis

---

**Product:** Dichloromethane  
**Cat:** CP-C8440  
**Lot:** EB681-X  
**Date of Manufacture:** 05/21/2021  
**Use Before Date:** 05/20/2024

Contains Cyclohexene Preservative  
Meets ACS Specifications for General Use

<u>Test</u>	<u>Specification</u>		<u>Result</u>	<u>Units</u>
	<u>Min.</u>	<u>Max.</u>		
Appearance	Clear		Clear	
Water by Karl Fischer Titration		0.02	0.003	%
Residue after Evaporation		0.002	<0.0001	%
Assay by GC Analysis	99.5		99.99	%
Color (APHA)		10	<10	
Titration Acid		0.3	0.06	µeq/g
Free Halogens	Passes test		Passes test	

Quality Control Approval

LIMS Sample No.: AM01631  
5/24/2021