



Certificate of Analysis

Iodine (Iodine-Iodide), 0.100 Normal (N/10)

Lot Number: 4210M24

Product Number: CP-1265

Manufacture Date: 10/27/2022

Expiration Date: 10/20/2023

The certified value for this product is confirmed in independent testing by a second qualified chemist.

To meet USP specifications for Iodine, Tenth-Normal (0.1 N) Volumetric Solution (VS) and Test Solution (TS) add 3-drops of Hydrochloric acid per Liter of R4000.

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Iodide	7681-11-0	ACS
Iodine	7553-56-2	ACS

Test	Specification	Result	NIST SRM#
Appearance	Dark brown liquid	Passed	
Assay (vs. Sodium Thiosulfate/Starch)	0.0998-0.1002 N at 20°C	0.1001 N at 20°C	136

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
CP-C1265	4 L amber glass	12 months

Recommended Storage: 15°C - 30°C (59°F - 86°F)

Paul Brandon

Production Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."