



39103 Warren Road
Westland, MI 48185
1.734.884.4773 Chempurebrand.com

CERTIFICATE OF ANALYSIS

Product: Water, Deionized, ASTM Type II
Date of Manufacture: 3/16/2023
Expiration Date: September 2024
Intended Use: General Laboratory Reagent/Chemical.
Recommended Storage: 15 to 25 °C (59 to 77 °F)
Product Code: CP-E0204C, CP-C0012P
Lot Number: 8303375
CAS Number: 7732-18-5
Country of Origin: United States of America

Test	Specification	Result
Appearance	Clear, colorless	Clear, colorless
Chloride (Cl)	<= 5.00 ppb	5.00ppb
Conductivity at 25°C	<= 1.00 µS/cm	1.00µS/cm
Heavy Metals (as Pb)	<= 0.01 ppm	0.01ppm
Nitrate (NO ₃)	<= 400.00 ppb	400.00ppb
Organic Carbon (TOC)	<= 50 ppb	50ppb
Oxidizable Substances (Permanganate Retention)	To pass test	Passes
Phosphate (PO ₄)	<= 1,000.00 ppb	1,000.00ppb
Silica (SiO ₂)	<= 3.0 ppb	3.0ppb
Sodium (Na)	<= 5.00 ppb	5.00ppb
Sulfate (SO ₄)	<= 1,000.00 ppb	100.00ppb

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Quality Control, Certifying Officer

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials**Contents of Certificates and Labels".