



39103 Warren Road  
 Westland, MI 48185  
 1.734.884.4773 Chempurebrand.com

## CERTIFICATE OF ANALYSIS

<i>Product:</i> <b>Buffer, Reference Standard, pH = 7.00 (Yellow)</b>	<i>Date of Manufacture:</i> 8/8/2023
	<i>Expiration Date:</i> August 2025
<i>Intended Use:</i> Calibration of pH Meters.	<i>Recommended Storage:</i> 15 to 25 °C (59 to 77 °F)
<i>Product Code:</i> CP-E1238, CP-C1238P, CP-A1238P J2156000	<i>Lot Number:</i> <b>8308158</b>
<i>CAS Number:</i> 7732-18-5;7558-79-4;7778-77-0;26628-22-8	<i>Country of Origin:</i> United States of America

This solution is traceable to the NIST Standard Reference Materials (SRMs) Potassium Dihydrogen Phosphate (186-I), Disodium Hydrogen Phosphate (186-II), Sodium Bicarbonate (191-I) and Sodium Carbonate (191-II).

Test	Specification	Result
------	---------------	--------

pH at 25°C	6.99 - 7.01	7.00
------------	-------------	------

This buffer is tested in our Charlotte, NC laboratory under ISO|IEC 17025 accreditation (PJLA Accreditation 64539) and is standardized against National Institute of Standards and Technology (NIST) Traceable Buffer Solutions prepared according to exact NIST instructions. The solution is checked and verified using two independently calibrated high precision pH meters. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture and testing according to validated methods. Batch records document raw material traceability as well as production and testing history for each lot manufactured.

Quality Control, Certifying Officer

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials\*\*Contents of Certificates and Labels".