



Certificate of Analysis

Product Name: TISAB II

Product Code: CP-E2613P

Manufactured: Aug-21-2024

Lot # 4GH1125

Expiration: Aug - 2026

Analytical Results		
Item	Specifications	
Description		
pH	5.3 – 5.5	5.33
Slope	54 – 60 mV	58.7 mV

E = 20L

Disclaimer: To the best of our knowledge, the information contained herein is accurate and reliable as of the date of publication. However, nothing herein shall constitute any express or implied warranty of merchantability or fitness for a particular purpose. It is the customer's responsibility to inspect and test our products in order to satisfy itself as to the suitability of the products for the customer's particular purpose.