



...Pure Quality

Certificate of Analysis

Product: ACS/HPLC Acetonitrile
Cat: CP-C0157
Lot: EK557
Description: ACS/ HPLC Acetonitrile, ACS reagent for general use, for HPLC, Spectrophotometry, USP spec.
Date of Manufacture: 2025-May-29
Use Before Date: 2028-May-28

Meets ACS / USP Specifications

<u>Test</u>	<u>Specification</u>		<u>Result</u>	<u>Units</u>
	<u>Min.</u>	<u>Max.</u>		
Appearance	Clear		Clear	
Water by Karl Fisher Titration		0.05	0.001	%
UV Cutoff		190	188	nm
Refractive Index (20°C)	1.3434	1.3446	1.3438	
Residue		3	<1	mg/L
GC Analysis	99.9		99.99	%
Color (APHA)		10	<10	
Titration Acid		8.0	0.20	µeq/g
Titration Base		0.6	0.01	µeq/g
HPLC Gradient @ 254 nm		0.005	<0.0050	AU
UV Absorbance @ 200 nm		0.070	0.0168	AU
UV Absorbance @ 210 nm		0.040	0.0126	AU
UV Absorbance @ 225 nm		0.020	0.0052	AU
UV Absorbance @ 250 nm		0.005	<0.0001	AU
UV Absorbance @ 350 nm		0.005	<0.0001	AU

Quality Control Approval

LIMS Sample No.: AQ01340
5/29/2025