



...Pure Quality

Certificate of Analysis

Product: Hexane (>95%)
Cat: CP-C1496
Lot: EK619
Description: For GC, for pesticide residue analysis, non-spectrophotometric grade
Date of Manufacture: 2025-June-11
Use Before Date: 2028-June-10

<u>Test</u>	<u>Specification</u>		<u>Result</u>	<u>Units</u>
	<u>Min.</u>	<u>Max.</u>		
Water by Karl Fisher Titration		0.010	0.003	%
Residue		1	<0.5	mg/L
GC Analysis (total C6 isomers)	99.9		>99.99	%
GC Analysis (n-Hexane)	95.0		99.33	%
Electron Capture GC		10	<10	ng/L
Flame Ionization GC		5	<5	ppb
Color (APHA)		10	<10	
Thiophene	Passes test		Passes test	

Quality Control Approval

LIMS Sample No.: AQ01452
6/11/2025