

## CERTIFICATE OF ANALYSIS

*Product:* **Buffer, Reference Standard, pH = 4.00 (Red)**  
*Date of Manufacture:* 12/10/24  
*Expiration Date:* January 2027  
*Intended Use:* Calibration of pH Meters.  
*Recommended Storage:* 15 to 25 °C (59 to 77 °F)  
*Product Code:* CP-E1228, CP-C1228P, CP-A1228P  
*Lot Number:* **8412104**  
*Country of Origin:* United States of America

This solution is traceable to the NIST Standard Reference Materials (SRMs) Potassium Dihydrogen Phosphate (186-I), Disodium Hydrogen Phosphate (186-II) and Potassium Hydrogen Phthalate (185).

Test	Specification	Result
------	---------------	--------

pH at 25°C	3.99 - 4.01	4.00
------------	-------------	------

Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.



Quality Control, Certifying Officer

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials\*\*Contents of Certificates and Labels".