



...Pure Quality

---

## Certificate of Analysis

---

**Product:** Hexane (>95%)  
**Cat:** CP-C1496  
**Lot:** EM206  
**Description:** For GC, for pesticide residue analysis, non-spectrophotometric grade  
**Date of Manufacture:** 2026-May-05  
**Use Before Date:** 2029-May-04

<u>Test</u>	<u>Specification</u>		<u>Result</u>	<u>Units</u>
	<u>Min.</u>	<u>Max.</u>		
Water by Karl Fisher Titration		0.010	0.001	%
Residue		1	<0.5	mg/L
GC Analysis (total C6 isomers)	99.9		>99.99	%
GC Analysis (n-Hexane)	95.0		99.35	%
Electron Capture GC		10	<10	ng/L
Flame Ionization GC		5	<5	ppb
Color (APHA)		10	<10	
Thiophene	Passes test		Passes test	

Quality Control Approval

LIMS Sample No.: AR01117  
5/6/2026